

血细胞分析仪用质控品 Z-3D Z-3D Control

质控品靶值表及范围
Zybio Control target value and range

LOT H201001
2021-02-10

仪器型号 Instrument Model	参数 Parameter	低值 Low	中值 Normal	高值 High	
Z3、Z3 CRP Z30、Z30 CRP Z31、Z31 CRP	WBC	$10^9/L$	2.06 ± 0.41	8.29 ± 1.24	22.39 ± 3.36
	Lym#	$10^9/L$	1.21 ± 0.24	2.71 ± 0.54	3.37 ± 1.35
	Mid#	$10^9/L$	0.16 ± 0.13	0.46 ± 0.37	0.89 ± 0.71
	Gran#	$10^9/L$	0.70 ± 0.21	5.13 ± 0.77	18.13 ± 1.81
	Lym%	%	58.6 ± 11.7	32.6 ± 13.1	15.1 ± 6.0
	Mid%	%	7.6 ± 6.1	5.6 ± 4.4	4.0 ± 3.2
	Gran%	%	33.8 ± 10.1	61.8 ± 12.4	80.9 ± 8.1
	RBC	$10^{12}/L$	2.43 ± 0.15	4.74 ± 0.28	5.86 ± 0.35
	HGB	g/L	60 ± 4	143 ± 7	198 ± 8
	HCT	%	17.2 ± 1.6	41.2 ± 3.7	55.5 ± 5.0
	MCV	fL	70.9 ± 5.0	86.9 ± 6.1	94.8 ± 6.6
	MCH	p/g	24.8 ± 1.7	30.3 ± 2.1	33.7 ± 2.4
	MCHC	g/L	350 ± 28	349 ± 28	356 ± 28
	RDW-CV	%	18.3 ± 3.0	17.1 ± 3.0	17.4 ± 3.0
	RDW-SD	fL	43.3 ± 6.0	48.5 ± 6.0	54.1 ± 6.0
	PLT	$10^9/L$	65 ± 13	281 ± 42	585 ± 88
	MPV	fL	9.1 ± 3.0	10.0 ± 3.0	10.1 ± 3.0
	PDW	fL	16.0 ± 3.0	16.4 ± 3.0	16.6 ± 3.0
	PCT	%	0.059 ± 0.050	0.282 ± 0.100	0.588 ± 0.200
	P-LCC	$10^9/L$	13 ± 10	74 ± 22	154 ± 30
P-LCR	%	20.3 ± 9.0	26.2 ± 9.0	26.4 ± 9.0	

注意事项:在使用血细胞分析仪用质控品Z-3D之前, 请阅读包装说明书。
血细胞分析仪用质控品Z-3D开瓶有效期为14天。

Notes: Before using, refer to the instruction sheet for mixing directions.
Z-3D Control has 14-day open vial stability.

Zybio Inc.

住所: 重庆市大渡口区建桥工业园C区太康路6号30栋第1-4层 邮编: 400082 联系方式: 023-68190901 E-mail: zy@zybio.com
网址: <http://www.zybio.com>
Add: Floor 1 to Floor 4, Building 30, No.6 of Taikang Road, Block C of Jianqiao Industrial Park, Dadukou District, Chongqing, China 400082
Tel: +86 (0) 23 6865 5509 E-mail: info@zybio.com Web: www.zybio.com

